



NATA ACCREDITED LABORATORY

National Association of Testing Authorities, Australia

(ABN 59 004 379 748)

has accredited

LABGEAR AUSTRALIA PTY LTD
LabGear Australia

following demonstration of its technical competence to operate in accordance with

ISO/IEC 17025

This facility is accredited in the field of

MECHANICAL TESTING

for the tests shown on the *Scope of Accreditation* issued by NATA

Jennifer Evans

Chief Executive Officer

Date of issue: 16 March 2016

Date of accreditation: 11 March 2016

Accreditation number: 19746

Scope of Accreditation



ACCREDITATION NO: 19746

LABGEAR AUSTRALIA PTY LTD

LabGear Australia
30 Finchley Street
MILTON QLD 4064

CONTACT: Ms Clare Williams
PHONE: 07 3368 1112 FAX: 07 3369 9885 MOBILE: 0418 924 113
EMAIL: clare.williams@labgearaustralia.com.au
WEB SITE: www.labgearaustralia.com.au

FACILITIES: Public testing service

This facility complies with the requirements of ISO/IEC 17025:2005

13.69 Controlled environments

.01 Clean rooms and workstations

Air velocity, uniformity of air velocity, work zone integrity,
integrity of terminally mounted HEPA filter installations,
illuminance,
by the methods of:
AS 1807.1; .5; .6; .15
For the purposes of determining compliance with
AS2252.1 cl. 5.2, 5.3.2, AS 2252.2 cl. 5.2.1, 5.2.3, 5.2.4, 5.3.4,
AS 2252.6 cl 3.2.1, 3.3.2 and AS 2567 cl 5.2.1, 5.2.2, 5.2.3, 5.2.4, 5.3.2

.02 Biological safety cabinets

Filter pre-scan, air velocity, uniformity of air velocity, work zone integrity,
integrity of terminally mounted HEPA filter installations,
illuminance, sound level, air barrier containment,
by the methods of
AS 1807.1; .5; .6; .15; .20; .21; .22; .26
For the purposes of determining compliance with
AS2252.1 cl. 5.2, 5.3.1, 5.3.2, AS 2252.2 cl. 5.2.1, 5.2.2, 5.2.3, 5.2.4, 5.3.3, 5.3.4,
AS 2252.6 cl 3.2.1, 3.3.2, 3.3.3 and AS 2567 cl 5.2.1, 5.2.2, 5.2.3, 5.2.4, 5.3.1, 5.3.2

Accreditation No: 19746
(Scope Last Changed 15/03/16)
